

Collaborative Research: NRI: INT: Customizable Lower-Limb Wearable Robot using Soft-Wearable Sensor to Assist Occupational Workers

Myunghee Kim (PI, UIC), Heejin Jeong (Co-PI, UIC), W. Hong Yeo (PI, Georgia Tech)



NSF-NRI, NIOSH

Award ID#: 2024863, 2024742

Goal: Personalization in lower-limb assistive wearable robots using human-in-the-loop (HIL) optimization to reduce the physical effort in intensive activities, thereby reducing injury.

Challenges

- 1) Inaccurate, slow physical effort estimation.
- 2) Error-prone, burdensome biofeedback sensors.
- 3) Evaluation for intensive activities.

Aim 1. Fast HIL optimization

- The importance of personalized assistance for squatting was tested
 - Developed fast model-free, phase-plane based metabolic cost estimation method (Fig. 1B) [1].
 - Implemented a squat assistance controller in a robotic ankle exoskeleton (Fig. 1AD) [2].
 - Conducted human-in-the-loop optimization (Fig. 1)
 - The optimized assistance significantly reduced the metabolic cost of squatting (Fig. 2)




Fig. 1 Human-in-the-loop (HIL) optimization for squatting.




Fig. 2 The optimized assistance reduced the metabolic cost of squatting (left). Each participant presented a different optimal assistance parameter set (right).

ECG metric can be an alternative measure




Fig. 3 RMSSD (left), an ECG measure, presented a high correlation to the metabolic cost of squatting (right).

Metric using EMG can be an alternative

- The optimized assistance reduced activities of muscle synergies (muscle coordination patterns) [3]




Fig. 4 Muscle coordination and activation patterns during squatting for the no-device and optimized assistance conditions.

A pressure sensor can be a potential measure




Fig. 5 Personalized assistance appears to change CoP variability [4].

Aim 2: Soft wearable electronics



Fig. 5 Bulky Commercial calorimetric devices and soft wearable electronics (red circle). [5]

Aim 3: Seamless integration & evaluation




Fig. 6 AFO and biopatch with calorimetry mask (left) and bio-signal from soft wearable electronics (right) [5].

Impact: Science & Technology, Applications / Education

- Efficient physical effort estimation for customizing assistance to assist workers.
- Integrates research & education via projects with improved infrastructure
- Wireless, soft wearable electronics, applicable for diagnosis and monitoring.

References

- [1] Kantharaju and Kim, in review
[2] Kantharaju et. al., in review
[3] Jeong et.al., in preparation
[4] Ramadurai et.al., in review
[5] Kim et. al, in preparation.